
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRU00.04
		Att.05
		RU91B00.01

TECHNOZYM® ADAMTS-13 Activity ELISA 96T.

REF	5450701
LOT	RU91B00.01
	2020-07-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
2x	6mL	ADAMTS-13 Activity vWF73 Substrate	900	2020-07-31
1x	30mL	ADAMTS-13 Activity Reaction Buffer	900	2020-07-31
1x	12mL	ADAMTS-13 Activity HRP Conj Antibody	900	2020-07-31
1x	12mL	ADAMTS-13 Activity Color Reagent	900	2020-07-31
1x	53mL	ADAMTS-13 Activity Wash Buffer Conc	900	2020-07-31
1x	12mL	ADAMTS-13 Activity Stop Solution	900	2020-07-31
1x	n.a.	ADAMTS-13 Activity Plate 96T.	900	2020-07-31
1x	n.a.	ADAMTS-13 Activity sample dilution MTP	900	2020-07-31
3x	n.a.	ADAMTS-13 Activity Plate Seal	900	2020-07-31
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 1	1UT91B0	2021-01-31
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 2	2UT91B0	2021-01-31
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 3	3UT91B0	2021-01-31
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 4	4UT91B0	2021-01-31
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 6	6UT91B0	2021-01-31
1x	0.5mL	UY ADAMTS-13 Activity High Control	RUY91B0	2021-01-31
1x	0.5mL	UZ ADAMTS-13 Activity Low Control	RUZ91B0	2021-01-31
1x	0.5mL	UT ADAMTS-13 Activity Calibrator 5	5UT91B0	2021-01-31

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	deltaOD Cal1-Cal6	≥ 0.7	1.5	OK
Kontrollwiederfindung Recovery of controls	ADAMTS-13 Activity High Control LOT RUY91B0	0.50 - 0.84 IU/mL	0.70 IU/mL	OK
	ADAMTS-13 Activity Low Control LOT RUZ91B0	0.13 - 0.29 IU/mL	0.24 IU/mL	OK



**ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS**

DQRU00.04

Att.05

RU91B00.01

Durchgeföhrt von
Prepared by

Michael KAFKA 

21. FEB. 2019

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

21. FEB. 2019

Datum / Date: