



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ9600.01

Ceveron® Cleaning Solution 3% 25 mL

REF 9820300

LOT 9683B00.01

🕒 2019-12-31

Analytische Verfahren Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	11.5 – 12.5	12.4	OK
Volumen Volume	25.0 – 26.5 mL	26.0	OK
Sterilität Sterility	0 CFU/mL	0	OK

Verwendbarkeit, Fibrinogen, Clauss method Ceveron® alpha Usability, Fibrinogen, Clauss method Ceveron® alpha	CCN LOT 1P71B01	3.17 – 4.75 g/L	4.44	OK
	CCA LOT 3P71B01	1.02 – 1.54 g/L	1.23	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Eva Maria HAGER

21. JAN. 2019

Datum / Date:

Freigabe von  
Approved by

Suzana GRESAK PODAKAR

22. JAN. 2019

Datum / Date: