

# CERTIFICATE OF ANALYSIS

**CHROMOGENIX**

Instrumentation  
Laboratory  
A Werfen Company

PRODUCT NAME:	CoaTest SP Factor FVIII Kit	EXPIRATION DATE:	2020-09
PRODUCT NUMBER:	82408663		
LOT NUMBER :	N1081014		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
<b>High Range</b>					
Slope	0.0077	0.0112	--	0.0088	PASS
Intercept	--	0.250	--	0.145	PASS
Standard Curve (r)	0.99	--	--	0.997	PASS
<b>Low Range</b>					
Slope	0.020	0.060	--	0.040	PASS
Intercept	--	0.406	--	0.155	PASS
Standard Curve (r)	0.99	--	--	0.999	PASS

PREPARED BY: Nehal Patel *N. m. Patel* DATE: 10/23/18  
PRINT SIGNATURE

TITLE Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza* DATE: 10/23/18 *10/23/18*  
PRINT SIGNATURE

TITLE Senior QC Analyst