



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ6S0A.01

Siron LS (aPTT liquid) 10x4 mL

REF 5035107

LOT 6S84BA0.01

⌚ 2020-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.6	7.2	OK
Sterilität sterility		0 CFU/mL	0	OK
Sensitivität Lupus sensitivity lupus	Lupus Inhibitor Plasma	≥ 1.5	3.7	OK
aPTT Verwendbarkeit aPTT Usability Ceveron® alpha	CCN LOT 1P71B01	23.0 – 28.4 sec	24.5	OK
	CCA LOT 3P71B01	46.8 – 57.8 sec	51.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Eva Maria HAGER

18. OKT. 2018  
Datum / Date:

Freigabe von  
Approved by

Suzana GRESAK PODAKAR

18. OKT. 2018  
Datum / Date: