



	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2600.01
-----------------------------------------------------------------------------------	-------------------------------------------------------	-----------


TECHNOVIEW® Arixtra CAL 5x1 mL

REF 5090010
LOT 2691B00.01
 2020-12-31

Inhalt Content	Reagenz Reagent	LOT	
1x1 mL	TECHNOVIEW Arixtra Cal 1	2181B00	2020-12-31
1x1 mL	TECHNOVIEW Arixtra Cal 2	2281B00	2020-12-31
1x1 mL	TECHNOVIEW Arixtra Cal 3	2381B00	2020-12-31
1x 1mL	TECHNOVIEW Arixtra Cal 4	2481B00	2020-12-31
1x1 mL	TECHNOVIEW Arixtra Cal 5	2581B00	2020-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Arixtra Cont low LOT 3U71B00	0.32 – 0.59 µg/mL	0.50	OK
	TECHNOVIEW Arixtra Cont high LOT 3V71B00	0.59 – 0.98 µg/mL	0.78	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Stefan STEINER

22. MRZ. 2019

Datum / Date:

**Freigabe von
Approved by**


Suzana GRESAK PODAKAR

22. MRZ. 2019

Datum / Date: