
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3F00.01
-----------------------------------------------------------------------------------	-------------------------------------------------------	-----------

TECHNOVIEW® LMW Low CONT 5x1mL

REF 5090042
LOT 3F81B00.01
 2020-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		0.25 – 0.47 IU/mL	0.36	OK
Verwendbarkeit Usability Ceveron © Alpha	Anti-Xa LMWH	0.25 – 0.47 IU/mL	0.41	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


**Durchführung von
Prepared by**



Christoph ENNE

22. JUNI 2018
Datum / Date:

**Freigabe von
Approved by**



Kaydinn NASUFI

22. JUNI 2018
Datum / Date: