


| | | |
|---|---|-----------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQ3H00.02 |
|---|---|-----------|

TECHNOVIEW® LMW CONT H 5x1ml

REF 5090046

LOT 3H81B00.01

 2020-04-30


| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|---|----------------------------------|----------------------------------|----|
| pH | | 7.2 – 7.9 | 7.7 | OK |
| Virologie Virology | HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab | Negative Negative Negative | Negative Negative Negative | OK |
| Richtwert Expected value Ceveron® alpha | Anti-Xa | 0.83 – 1.39 IU/mL | | |
| Verwendbarkeit Usability Ceveron® alpha | Anti-Xa | 0.83 – 1.39 IU/mL | 1.11 | OK |
| Etiketten und Barcode Überprüfung Label and bar code Verification | | | OK | |

**Durchführung von
Prepared by**


Christoph ENNE

27. JUNI 2018
Datum / Date:

**Freigabe von
Approved by**


Lavdym NASIFI

27. JUNI 2018
Datum / Date: