



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0Y00.01

TECHNOVIEW UFH CAL 5x1mL

REF 5090070

LOT 0Y81B00.01

⌚ 2020-01-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1mL	TECHNOVIEW UFH Cal 1	5F13B04	2020-01-31
1x1mL	TECHNOVIEW UFH Cal 2	5G13B04	2020-01-31
1x1mL	TECHNOVIEW UFH Cal 3	5H13B04	2020-01-31
1x1mL	TECHNOVIEW UFH Cal 4	5J13B04	2020-01-31
1x1mL	TECHNOVIEW UFH Cal 5	5K13B04	2020-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron [®] alpha	TECHNOVIEW UFH Cont low LOT 3C71B00	0.18 – 0.34 IU/mL	0.26 OK
	TECHNOVIEW UFH Cont high LOT 3J71B00	0.38 – 0.63 IU/mL	0.51 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by

Christoph ENNE

19. JULI 2018

Datum / Date:

Freigabe von
Approved by

Eva Maria HAGER

19. JULI 2018

Datum / Date: