
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0J00.01
-----------------------------------------------------------------------------------	-------------------------------------------------------	-----------

TECHNOCHROM[®] C1-Inhibitor 30 T.

REF 5345003
LOT 0J91B00.01
 2020-09-30

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	Coagulation Control N for C1-INH	1P81BC1H	2020-10-31
1x1mL	Coagulation Control A for C1-INH	3P82BC1H	2020-11-30
1x0.953 mL	Coagulation Reference for C1-INH	1R81BC1H	2020-10-31
1x3 mL	C1 Esterase	7B81B00	2020-09-30
1x3 mL	Substrate f. C1-Esterase	8C72B00	2022-06-30
1x25 mL	C1-INH Probenpuffer A	9M81B00	2023-10-31
1x20 mL	C1-INH Reaktionspuffer B	9N81B00	2023-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron [®] alpha	CCN LOT 1P81BC1H	85.1 – 141.8 %	116.6	OK
	CCA LOT 3P82BC1H	35.7 – 59.5 %	44.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


Durchführung von
Prepared by


Eva Maria HAGER

27. FEB. 2019

Datum / Date:

Freigabe von
Approved by


Stefan STEINER

27. FEB. 2019

Datum / Date: