

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQxUT0.02
		Alt.21
		7UT82B0.01

TECHNOZYM® ADAMTS-13 Activity Control Set 2 x 0.5 mL

REF	5450763
LOT	7UT82B0.01
	2020-06-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5mL	ADAMTS-13 Activity High Control	RUY81B0	2020-06-30
1x	0.5mL	ADAMTS-13 Activity Low Control	RUZ81B0	2020-06-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Kontrollwiederfindung Recovery of controls	ADAMTS-13 Activity High Control LOT RUY81B0	0.53 - 0.89 IU/mL	0.74 IU/mL	OK
	ADAMTS-13 Activity Low Control LOT RUZ81B0	0.13 - 0.29 IU/mL	0.24 IU/mL	OK

Durchgeführt von
Prepared by

Fikret NASUFI *Fikret Nasufi*

11. SEP. 2018

Datum / Date:

Freigabe von
Approved by

Christoph Enne
Christoph ENNE

11. SEP. 2018

Datum / Date: