

CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation
Laboratory

A Wernick Company

PRODUCT NAME:	S-2765, 25 mg Substrate	EXPIRATION DATE:	2021-11
PRODUCT NUMBER:	82141339		
LOT:	N1181952		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION PASS/FAIL
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.8	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.2	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 11-27-18
 PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 11-28-18
 PRINT SIGNATURE

TITLE: QC Technical Supervisor