



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ...00AZ

TECHNOZYM[®] PAP Complex ELISA Kit 96T.

REF TC12060
LOT 120601B9.01
2019-09-30

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x	PAP Plate	RGY82B0	2020-04-30
1 x 0.3mL	anti PAP POX	RBK81B0	2020-03-31
2 x 20mL	Wash buffer Concentrate	RBE81B0	2020-04-30
1 x 20mL	Dilution Buffer	RAF91B0	2021-01-31
1 x 12mL	TMB Substrate	RKN82B0	2020-05-31
1 x 12mL	Stop solution	RKK81B0	2020-04-30
1 x 0.5mL	PAP-Calibrator 1	1BM61B1	2019-09-30
1 x 0.5mL	PAP-Calibrator 2	2BM61B1	2019-09-30
1 x 0.5mL	PAP-Calibrator 3	3BM61B1	2019-09-30
1 x 0.5mL	PAP-Calibrator 4	4BM61B1	2019-09-30
1 x 0.5mL	PAP-Calibrator 5	5BM61B1	2019-09-30
1 x 0.5mL	PAP-high Control	ABM61B1	2019-09-30
1 x 0.5mL	PAP-low Control	BBM61B1	2019-09-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	OD: 1 – 2.5	1.33	OK
	Calibrator 5	OD: < 0.2	0.02	OK
Verwendbarkeit Usability	High Control (ABM61B1)	98.9 – 164.9 ng/mL	129.56	OK
	Low Control (BBM61B1)	13.0 – 30.2 ng/mL	20.83	OK

Durchführung von
Prepared by

Michael KAFKA

05. FEB. 2019

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

05. FEB. 2019

Datum / Date: