
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7C00AZ
---	---	----------

TECHNOTHROMBIN® TGA RB 5x0.5mL

REF 5006209

LOT 7C91B00.01

 2021-01-31

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RB	7C91B00	2021-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z83B00	152 - 253	195.9	OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

09. APR. 2019
Datum / Date:

**Freigabe von
Approved by**


Eva Maria HAGER

10. APR. 2019
Datum / Date: