



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS


DQ1Z00AZ

TECHNOTHROMBIN[®] TGA Control High 5x1mL

REF 5006320

LOT 1Z83B00.01

 2019-11-30

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1Z83B00	2019-11-30

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC Low (nM Peak Thrombin) LOT 7D83B00	260 - 433 276.4 OK
	TGA Reagenz RC High (nM Peak Thrombin) LOT 7F81B00	410 - 684 472.9 OK

Durchführung von
Prepared by

Fikret NASUFI 

29. MRZ. 2019

Datum / Date:

Freigabe von
Approved by

Michael KAFKA 

29. MRZ. 2019

Datum / Date: