
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2800AZ
---	---	----------

TECHNOTHROMBIN[®] TGA Control Low 5x1mL

REF 5006330

LOT 2891B00.01

 2020-03-31

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control Low	2891B00	2020-03-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC Low (nM Peak Thrombin) LOT 7D83B00	150 - 250	216.8 OK
	TGA Reagenz RC High (nM Peak Thrombin) LOT 7F81B00	248 - 413	372.7 OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

29. MRZ. 2019
Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *[Signature]*

29. MRZ. 2019
Datum / Date: