

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ0J00.01
---	---	-----------

## TECHNOCHROM® C1-Inhibitor 30 T.

REF      5345003  
LOT      0J91B00.02  
      2020-09-30

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1x1mL	Coagulation Control N for C1-INH	1P81BC1H	2020-10-31
1x1mL	Coagulation Control A for C1-INH	3P82BC1H	2020-11-30
1x0.953 mL	Coagulation Reference for C1-INH	1R81BC1H	2020-10-31
1x3 mL	C1 Esterase	7B81B00	2020-09-30
1x3 mL	Substrate f. C1-Esterase	8C72B00	2022-06-30
1x25 mL	C1-INH Probenpuffer A	9M81B00	2023-10-31
1x20 mL	C1-INH Reaktionspuffer B	9N81B00	2023-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN <span style="border: 1px solid black; padding: 2px;">LOT</span> 1P81BC1H	85.1 – 141.8 %	107.7	OK
	CCA <span style="border: 1px solid black; padding: 2px;">LOT</span> 3P82BC1H	35.7 – 59.5 %	44.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
**Eva Maria HAGER**

11. APR. 2019  
Datum / Date:

**Freigabe von**  
**Approved by**

  
**Stefan STEINER**

11. APR. 2019  
Datum / Date: