

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Z00AZ.02
---	---	-------------

TECHNOCHROM® FXIII 3 x 3 mL

REF 5360010
LOT 0Z91B00.01
 2020-03-31

Inhalt	Reagenz	LOT	
3 x 3 mL	FXIII Activator Reagent	1901100/1	2020-03-31
3 x 3 mL	FXIII Detection Reagent	1901100/2	2020-03-31
3 x 3 mL	FXIII NADPH	1901100/3	2020-03-31
3 x 1 mL	FXIII Inhibitor Reagent	1901100/4	2020-03-31
1 x 6 mL	Stabilizer solution	1901100/5	2020-03-31

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P71B01	58.2 – 97.0 %
	Coagulation Control A Lot: 3P71B01	28.8 – 48.0 %

Durchführung von
Prepared by

Michael KAFKA 

15. APR. 2019

Datum / Date:

Freigabe von
Approved by


Eva Maria HAGER

15. APR. 2019

Datum / Date: