



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRA00AZ

TECHNOZYM® vWF:Ag ELISA Kit 96T

REF 5450201

LOT RA91B00.01

🕒 2020-05-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1 x	vWF:Ag Plate	RNF82B0	2020-08-31
1 x 0.3mL	Ag Conjugate	RNH81B0	2020-05-31
1x 80mL	Wash buffer Concentrate	RNA83B0	2020-10-31
1 x 90mL	Incubation buffer	RNB83B0	2020-10-31
1 x 12mL	TMB Substrate	RKN83B0	2020-12-31
1 x 12mL	Stop solution	RKK82B0	2020-08-31
1 x 0.5mL	vWF Calibrator 1	1ND71B1	2021-01-31
1 x 0.5mL	vWF Calibrator 2	2ND71B1	2021-01-31
1 x 0.5mL	vWF Calibrator 3	3ND71B1	2021-01-31
1 x 0.5mL	vWF Calibrator 4	4ND71B1	2021-01-31
1 x 0.5mL	vWF Calibrator 5	5ND71B1	2021-01-31
1 x 0.5mL	vWF High Control	AND71B1	2021-01-31
1 x 0.5mL	vWF Low Control	BND71B1	2021-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.42	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (AND71B1)	0.68 – 1.13 IU/mL	0.86	OK
	Low Control (BND71B1)	0.16 – 0.38 IU/mL	0.28	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

02. APR. 2019

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

02. APR. 2019

Datum / Date: