

CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation
Laboratory
A Werfen Company


PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE: 2020-10
PRODUCT NUMBER:	82266863	
LOT NUMBER:	N0394338	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		
Analysis with COATEST APC R:					
APT Time					
<u>CaCl₂</u>	27	44	seconds	44	PASS
<u>APC/CaCl₂</u>	--	101	seconds	86	PASS
APC Ratio	--	2.3	--	1.9	PASS

Analysis with COATEST APC R V:					
APT Time					
<u>CaCl₂</u>	27	40	seconds	36	PASS
APC Ratio	1.4	1.9	--	1.6	PASS


We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: D.Tencza 
PRINT SIGNATURE

DATE: 3/18/2019 3/18/19

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: D.Harz 
PRINT SIGNATURE

DATE: 3/18/2019 3/18/19

TITLE: Technical Supervisor