

CERTIFICATE OF ANALYSIS

CHROMOGENIX


**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2021-01
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N0193516	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.38	PASS
Intercept	---	0.12	--	0.05	PASS
r	0.990	---	--	0.999	PASS
LOW RANGE					
Slope	2.8	6.9	--	4.5	PASS
Intercept	---	0.21	--	0.09	PASS
r	0.990	---	--	0.999	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel *N.M. Patel*

PRINT SIGNATURE

DATE: 2-19-19
2/19/2019

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz*

PRINT SIGNATURE

DATE: 2/19/19
2/19/2019

TITLE: Technical Supervisor