


	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7D00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA RC Low 5x0.5mL

REF 5006212
LOT 7D91B00.01
 2021-04-30

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RC Low	7D91B00	2021-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z91B00	249 - 415	355.8	OK
	TGA Control Low (nM Peak Thrombin) LOT 2893B00	139 - 231	218.6	OK

Durchführung von
Prepared by

Fikret NASUFI 

15. MAI 2019
Datum / Date:

Freigabe von
Approved by

Michael KAFKA 

15. MAI 2019
Datum / Date: