
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ7F00AZ
---	---	----------

## TECHNOTHROMBIN<sup>®</sup> TGA RC High 5x0.5mL

REF            5006214

LOT            7F92B00.01

            2021-05-31

Inhalt Content	Reagenz Reagent	LOT	
5x0.5 mL	TGA RC High	7F92B00	2021-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z91B00	417 - 695	609.8	OK
	TGA Control Low (nM Peak Thrombin) LOT 2893B00	211 - 351	303.3	OK

**Durchführung von  
Prepared by**

Fikret NASUFI *Fikret Nasufi*

14. MAI 2019

Datum / Date:

**Freigabe von  
Approved by**

Michael KAFKA *[Signature]*

15. MAI 2019

Datum / Date: