

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ8F00AZ
---	---	----------

TECHNOTHROMBIN® TGA SUB 50x1.5mL

REF 5006230
LOT 8F91B00.01
 2021-04-30

Inhalt Content	Reagenz Reagent	LOT	
50x1.5mL	TGA Substrat	8F91B00	2021-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z83B00	TGA RC High LOT 7F81B00	410 - 684	593.8	OK
		TGA RC Low LOT 7D83B00	260 - 433	356.0	OK
	TGA Control Low (nM Peak Thrombin) LOT 2891B00	TGA RC High LOT 7F81B00	248 - 413	311.9	OK
		TGA RC Low LOT 7D83B00	150 - 250	196.3	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

09. MAI 2019

Datum / Date:

Freigabe von
Approved by

Stefan STEINER

09. MAI 2019

Datum / Date: