
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1Z00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA Control High 5x1mL

REF 5006320

LOT 1Z91B00.01

 2020-04-30

Inhalt Content	Reagenz Reagent	LOT	
5x1mL	TGA Control High	1Z91B00	2020-04-30

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Reagenz RC Low (nM Peak Thrombin) LOT 7D91B00	249 - 415	355.8 OK
	TGA Reagenz RC High (nM Peak Thrombin) LOT 7F92B00	417 - 695	609.8 OK

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

15. MAI 2019

Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *[Signature]*

15. MAI 2019

Datum / Date: