

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ3900.01
---	---	-----------

**TECHNOVIEW® Apixaban CONT Low 5x1ml**

REF      5090271  
LOT      3981B00.02  
      2020-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		64.2 – 119.2 ng/mL	91.7	OK
Verwendbarkeit Usability Ceveron © alpha	Anti-Xa Apixaban	64.2 – 119.2 ng/mL	108.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
Suzana GRESAK PODAKAR

13. MAI 2019  
Datum / Date:

**Freigabe von  
Approved by**

  
Stefan STEINER

14. MAI 2019  
Datum / Date: