



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ0F00.03

## F VIII Inhibitor Reagent Kit HCV neg. 2-4T.

REF 5152009

LOT 0F86B00.01

🕒 2020-03-31

Inhalt Content	Reagenz Reagent	LOT	🕒
2 x 2.63 mL	FVIII normal Plasma	1G81B00	2020-06-30
1 x 1 mL	FVIII Inhibitor Plasma	2K81B00	2020-05-31
1x 1 mL	FVIII Inhibitor Free Plasma	2M81B0D	2020-03-31
1 x 17 mL	Imidazol Puffer	9J84BA0	2023-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability FVIII clotting, Dapttin Ceveron® alpha	FVIII Inhibitor Plasma LOT 2K81B00	17.6 – 29.3 BE/mL	22.1	OK
	FVIII Inhibitor Free Plasma LOT 2M81B0D	< 0.4 BE/mL	< 0.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Eva Maria HAGER

14. NOV. 2018

Datum / Date:

Freigabe von  
Approved by

Christoph ENNE

14. NOV. 2018

Datum / Date: