
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ2K00.02
---	---	-----------

FVIII Inhibitorplasma HCV neg. 5x1mL

REF 5159010
LOT 2K81B00.02
 2020-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.4	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	FVIII Inhibitor clotting, Siron LS	14.9 – 24.8 BE/mL	20.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Eva Maria HAGER

04. JULI 2018
Datum / Date:

**Freigabe von
Approved by**


Suzana GRESAK PODAKAR

05. JULI 2018
Datum / Date: