	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1R00.01
---	---	-----------

Coagulation Reference, 5x1mL


REF 5220110

LOT 1R71B03.01

 2021-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron [®] alpha	Fibrinogen, Clauss method, diluted plasma CCN LOT 1P71B01	3.17 – 4.75 g/L	4.25	OK
	Fibrinogen, Clauss method, diluted plasma CCA LOT 3P71B01	1.02 – 1.54 g/L	1.20	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Lavdrim NASUFI

04. APR. 2019

Datum / Date:

**Freigabe von
Approved by**


Stefan STEINER

04. APR. 2019

Datum / Date: