

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1Y00AZ.02
---	---	-------------

Ceveron® TGA Control High 5x1mL

REF	5006322
LOT	1Y83B00.02
⌚	2019-11-30

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.4 – 7.8	7.5	OK
Verwendbarkeit: Usability: Ceveron® alpha	Peak Thrombin nM Ceveron® TGA RC Low LOT 7R92B	206 - 344	296.0 OK
	Peak Thrombin nM Ceveron® TGA RC High LOT 7Q91B	454 - 756	585.8 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

03. APR. 2019

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *Michael Kafka*

03. APR. 2019

Datum / Date: