

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ470D.02
		Alt.03
		4782B0D.01

F VII Mangelplasma immunads, 5x1mL

REF	5144015
LOT	4782B0D.01
	2020-03-31

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Cap color	blue	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Appearance before reconstitution	buff colored plug of lyophilized material	confirmed	OK
Appearance after reconstitution	buff colored liquid without visible particles	confirmed	OK	
Virologie / Virology	HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab The testing methods applied were FDA-approved or CE marked.	Negative Negative Negative	Negative Negative Negative	OK
Kontrollwiederfindung FVII am Ceveron® Alpha	Control N LOT 1P71B01	87.9 - 146.5 %	121.0 %	OK
Recovery of controls FVII on Ceveron® Alpha	Control A LOT 3P71B01	33.5 - 55.9 %	46.9 %	OK

Durchgeführt von
Prepared by


 Lavdriq NASUFI

20. AUG. 2018

Datum / Date:

Freigabe von
Approved by


 Christoph ENNE

20. AUG. 2018

Datum / Date: