

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1L00.01 LOT 1L82B00.02
---	---	------------------------------------

Platelet Poor Plasma 5x2 mL

REF 5343022
LOT 1L82B00.02
 2020-08-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.4 – 7.8	7.6	OK
Restfeuchte Residual moisture	≤1%	0.19	OK
OD _{320nm}	≤10	6.2	OK
Verwendbarkeit Usability Ceveron® alpha	LA Ratio: Technoclot® LA Screen, LOT 6Y81B00 Technoclot® LA Confirm, LOT 6X71B00	0.88 – 1.46	1.02 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by

Riham EL KAMMAR

El Kammar Riham

12. APR. 2019

Datum / Date:

Freigabe von
Approved by

Eva Maria Hager
Eva Maria HAGER

12. APR. 2019

Datum / Date: