



## Certificate of Analysis

**REAADS**  
**VWF Activity ELISA**

**Kit Lot: VA-110**  
**Plate Lot #: VAW-043**  
**Product #: 10826**  
**Expiration Date: 2020-08-01**

### Quality Control Results

| <b>VWF Activity</b>       |                    | <b>Testing Date: 2019-02-26</b> |                        |  |
|---------------------------|--------------------|---------------------------------|------------------------|--|
| <b>Attribute</b>          | <b>Value</b>       | <b>Specification</b>            | <b>Result by Assay</b> |  |
| Reference Plasma 100% OD  | 69 % of Normal     | ≥ 1.0                           | 1.602                  |  |
| Reagent Background O.D.   |                    | ≤ 0.100                         | 0.009                  |  |
| Reference Curve R-Squared |                    | ≥ 0.95                          | 0.98502                |  |
| <b>Control Recovery</b>   | <b>Label Range</b> |                                 | <b>Result by Assay</b> |  |
| Normal Control            | 73 - 109 %         |                                 | 83.2 %                 |  |
| Deficient Control         | 17 - 44 %          |                                 | 27.6 %                 |  |

This signed Certificate of Analysis affirms that the product has met the requirements of the relevant specifications as set by Corgenix, Inc.

Human sera used to prepare the controls included in this kit have been tested and shown to be nonreactive for HBsAG and negative for antibodies to HCV and HIV 1 and 2 by FDA required tests. However, all human blood derivatives, including patient samples, should be handled as though capable of transmitting disease.

This product is warranted to perform as described in the package insert. Corgenix, Inc. disclaims any implied warranty of merchantability or fitness for a particular use, and in no event shall Corgenix, Inc. be liable for consequential damage.

**Quality Assurance Approval By: Malissa Sailakham**

Signed:  2019-03-01  
Title: **Quality Assurance Supervisor**

Corgenix, Inc.  
11575 Main Street, Suite 400  
Broomfield, CO 80020  
Phone: (303) 457-4345  
Fax: (303) 457-4519