

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRL00AZ
---	---	----------

TECHNOZYM® t-PA Antigen EDTA ELISA Kit

REF TC12007

LOT RL92B00.01

 2020-08-31

Inhalt Content	Reagenz Reagent	LOT	
1x	GN-Anti t-PA Plate	RGN91B0	2021-01-31
1x0.3mL	KM-POX anti t-PA	RKM81B1	2021-01-31
1x80mL	Wash buffer Concentrate	RNA83B0	2020-10-31
1x90mL	Incubation buffer	RNB83B0	2020-10-31
1x12mL	TMB Substrat 12mL	RKN91B0	2021-01-31
1x12mL	Stop solution	RKK82B0	2020-08-31
1x20mL	Dilution buffer	RAD81B0	2020-12-31
1x0.5mL	Calibrator 1 t-PA	1AV82B0	2020-12-31
1x0.5mL	Calibrator 2 t-PA	2AV81B0	2020-12-31
1x0.5mL	Calibrator 3 t-PA	3AV81B0	2020-12-31
1x0.5mL	Calibrator 4 t-PA	4AV81B0	2020-12-31
1x0.5mL	Calibrator 5 t-PA	5AV81B0	2020-12-31
1x0.5mL	Control High t-PA	AAV82B0	2020-12-31
1x0.5mL	Control Low t-PA	BAV82B0	2020-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.78	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AAV82B0)	10.79 – 17.99 ng/mL	14.66	OK
	Low Control (BAV82B0)	2.01 – 4.69 ng/mL	3.30	OK

Durchführung von
Prepared by

Michael KAFKA 

01. APR. 2019

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

02. APR. 2019

Datum / Date: