
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRB00AZ
---	---	----------

TECHNOZYM® vWF:CBA ELISA Kit 96T

REF 5450301
LOT RB94B00.01
 2021-01-31

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	vWF:CBA Plate Collagen Typ III	RNC92B0	2021-04-30
1 x 0.3mL	CBA Conjugate	RNE91B0	2021-03-31
1x 80mL	Wash buffer Concentrate	RNA91B0	2021-03-31
1 x 90mL	Incubation buffer	RNB91B0	2021-03-31
1 x 12mL	TMB Substrate	RKN91B0	2021-01-31
1 x 12mL	Stop solution	RKK91B0	2021-02-28
1 x 0.5mL	vWF Calibrator 1	1ND71B1	2021-01-31
1 x 0.5mL	vWF Calibrator 2	2ND71B1	2021-01-31
1 x 0.5mL	vWF Calibrator 3	3ND71B1	2021-01-31
1 x 0.5mL	vWF Calibrator 4	4ND71B1	2021-01-31
1 x 0.5mL	vWF Calibrator 5	5ND71B1	2021-01-31
1 x 0.5mL	vWF High Control	AND71B1	2021-01-31
1 x 0.5mL	vWF Low Control	BND71B1	2021-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.83	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (AND71B1)	0.84 – 1.40 IU/mL	1.12	OK
	Low Control (BND71B1)	0.18 – 0.42 IU/mL	0.31	OK

**Durchführung von
Prepared by**

Michael KAFKA 

24. APR. 2019

Datum / Date:

**Freigabe von
Approved by**

Fikret NASUFI 

25. APR. 2019

Datum / Date: