



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM<sup>®</sup> ADAMTS-13 INH ELISA Kit 48T

REF 5450451

LOT RP97B0S.02

⌚ 2021-01-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 6	Adamts13 INH Plate	RPS92BS	2021-01-31
1 x 0.3mL	Adamts13 INH Conjugate	RPV91B0	2021-01-31
1x 80mL	Wash buffer Concentrate	RNA91B0	2021-03-31
1 x 90mL	Incubation buffer	RNB91B0	2021-03-31
1 x 12mL	TMB Substrate	RKN91B0	2021-01-31
1 x 12mL	Stop solution	RKK91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH High Control	RPY91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Low Control	RPZ91B0	2021-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.64	OK
	Calibrator 5	< 0.1	0.00	OK
Verwendbarkeit Usability	High Control (RPY91B0)	56.55 – 94.25 U/mL	79.75	OK
	Low Control (RPZ91B0)	15.64 – 36.48 U/mL	25.67	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

18. JUNI 2019

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *[Signature]*

18. JUNI 2019

Datum / Date: