



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRP00AZ

TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF 5450451

LOT RP9AB0S.01

🕒 2021-01-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1 x 6	Adamts13 INH Plate	RPS98BS	2021-06-30
1 x 0.3mL	Adamts13 INH Conjugate	RPV91B0	2021-01-31
1x 80mL	Wash buffer Concentrate	RNA91B0	2021-03-31
1 x 90mL	Incubation buffer	RNB91B0	2021-03-31
1 x 12mL	TMB Substrate	RKN91B0	2021-01-31
1 x 12mL	Stop solution	RKK91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH High Control	RPY91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Low Control	RPZ91B0	2021-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.8	OK
	Calibrator 5	< 0.1	0.0	OK
Verwendbarkeit Usability	High Control (RPY91B0)	56.55 – 94.25 U/mL	81.91	OK
	Low Control (RPZ91B0)	15.64 – 36.48 U/mL	29.27	OK

Durchführung von
Prepared by

Michael KAFKA

18. JULI 2019

Datum / Date:

Freigabe von
Approved by

Eva-Maria Hager
Eva Maria HAGER

18. JULI 2019

Datum / Date: