CERTIFICATE OF ANALYSIS

CHROMOGENIX Instrumentation Laboratory
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE:	2021-02
PRODUCT NUMBER :	82258563		
LOT NUMBER:	N0293938		

TEST DESCRIPTION		SPECIFICATIONS			RESULTS	PASS/FAIL
		MIN	MAX	UNITS		
HIGH RANGE						
	Slope	0.33	0.57	--	0.45	PASS
	Intercept	---	0.12	--	0.06	PASS
	r	0.990	---	--	0.998	PASS
LOW RANGE						
	Slope	2.8	6.9	--	5.4	PASS
	Intercept	---	0.21	--	0.11	PASS
	r	0.990	---	--	1.000	PASS

We hereby certify that the above product(s):
· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
· meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Maria Oettinger (PRINT) Maria Oettinger (SIGNATURE) DATE: 3-20-19

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz (PRINT) Debra Lee Harz (SIGNATURE) DATE: 3-21-19

TITLE: QC Technical Supervisor

QCM0082258563/49730503-1, Rev 02

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000