

CERTIFICATE OF ANALYSIS



PRODUCT NAME:	CoaTest SP4 FVIII Kit	EXPIRATION DATE:	2021-02
PRODUCT NUMBER:	82409463		
LOT NUMBER:	N0293940		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION	
	MIN	MAX	UNITS			PASS/FAIL
Coatest SP4 Factor VIII Kit (82409463)						
High Range	Slope	0.0077	0.0112	--	0.0088	PASS
	Intercept	--	0.250	--	0.109	PASS
	Standard Curve (r)	0.99	--	--	1.000	PASS
Low Range	Slope	0.020	0.060	--	0.043	PASS
	Intercept	--	0.406	--	0.181	PASS
	Standard Curve (r)	0.99	--	--	0.998	PASS

KIT COMPONENTS:

DESCRIPTION	LOT NUMBER	VOLUME
COATEST SP FVIII CACL2	N0293934	6 mL
COATEST SP FVIII FACTOR IXa + FACTOR X	N0293835	3 mL
COATEST SP FVIII S-2765 + I-2581	N0293932	12 mL
COATEST SP FVIII BUFFER	N0394624	20 mL
COATEST SP FVIII PHOSPHOLIPID	N0293831	2 mL

PREPARED BY: Nehal Patel *N-m. Patel*
 PRINT SIGNATURE

DATE: 4/10/2019 *4-10-19*

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza*
 PRINT SIGNATURE

DATE: 4/10/2019 *4/10/19*

TITLE: Senior QC Analyst