
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ7F00AZ
---	---	----------

TECHNOTHROMBIN[®] TGA RC High 50x0.5mL

REF 5006216

LOT 7F92B00.02

 2021-05-31

Inhalt Content	Reagenz Reagent	LOT	
50x0.5 mL	TGA RC High	7F92B00	2021-05-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z91B00	417 - 695	630.6 OK
	TGA Control Low (nM Peak Thrombin) LOT 2893B00	211 - 351	343.0 OK

**Durchführung von
Prepared by**

Fikret NASUFI 

12. AUG. 2019

Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA 

12. AUG. 2019

Datum / Date: