


|   |   |          |
|---|---|----------|
|  | <b>ANALYSENZERTIFIKAT<br/>CERTIFICATE OF ANALYSIS</b> | DQ7G00AZ |
|---|---|----------|

## TECHNOTHROMBIN<sup>®</sup> TGA RD Reagenz 5x1.5mL

REF            5006220

LOT            7G91B00.01

            2021-04-30

| Inhalt<br>Content | Reagenz<br>Reagent | LOT     |  |
|-------------------|--------------------|---------|---|
| 5x1.5 mL          | TGA RD Reagenz     | 7G91B00 | 2021-04-30  |

| Analytische Prozedur<br>Analytical Procedure  |   | Anforderung<br>Requirement | Ergebnis<br>Result |    |
|---|---|----------------------------|--------------------|----|
| Kontrollwiederfindung<br>Recovery of controls | TGA Control High<br>(nM Peak Thrombin)<br>LOT 1Z83B00 | 428 - 713                  | 600                | OK |
|   | TGA Control Low<br>(nM Peak Thrombin)<br>LOT 2891B00  | 431 – 719                  | 685                | OK |

**Durchführung von  
Prepared by**

Michael KAFKA 

18. JULI 2019

Datum / Date:

**Freigabe von  
Approved by**

  
Stefan STEINER

19. JULI 2019

Datum / Date: