



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ9P00AZ

TECHNOTHROMBIN[®] TGA BUF 5x3 mL

REF 5006360

LOT 9P91B00.02

2021-05-31

Inhalt Content	Reagenz Reagent	LOT	
5 x 3 mL	TGA Buffer	9P91B00	2021-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Verwendbarkeit / usability Gerät / Analyzer: Biotek Reader: 197091	Regression der Bezugskurve / regression of the calibration curve	≥ 0.99	0.99	OK	
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z83B00	TGA RC High LOT 7F92B00	410 – 684	601.4	OK
Kontrollwiederfindung Recovery of controls	TGA Control High (nM Peak Thrombin) LOT 1Z83B00	TGA RC Low LOT 7D91B00	260 - 433	372.4	OK

Durchführung von
Prepared by

Michael KAFKA

15. JULI 2019

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

15. JULI 2019

Datum / Date: