



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3U00.01

TECHNOVIEW Arixtra Control L 6x1mL

REF 5090012

LOT 3U81B00.01

🕒 2020-04-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Richtwert Expected value Ceveron® alpha	Anti Xa Arixtra	0.30 – 0.56 µg/mL	0.43	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa Arixtra	0.30 - 0.56 µg/mL	0.35	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Suzana GRESAK PODAKAR

15. JUNI 2018

Datum / Date:

Freigabe von
Approved by

Stefan STEINER

15. JUNI 2018

Datum / Date: