



	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Z00.02
---	---	-----------

TECHNOCHROM® FXIII 3 x 3 mL

REF	5360010
LOT	0Z92B00.02
	2020-06-30

Inhalt	Reagenz	LOT	
3 x 3 mL	FXIII Activator Reagent	1902100/1	2020-06-30
3 x 3 mL	FXIII Detection Reagent	1902100/2	2020-06-30
3 x 3 mL	FXIII NADPH	1902100/3	2020-06-30
3 x 1 mL	FXIII Inhibitor Reagent	1902100/4	2020-06-30
1 x 6 mL	Stabilizer solution	1902100/5	2020-06-30

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P71B01	58.2 – 97.0 % 80.9
	Coagulation Control A Lot: 3P71B01	28.8 – 48.0 % 38.1

Durchführung von
Prepared by


Eva Maria HAGER

26. JUNI 2019
Datum / Date:

Freigabe von
Approved by

Michael KAFKA


26. JUNI 2019
Datum / Date: