



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRR00AZ

TECHNOZYM® ADAMTS-13 Antigen ELISA Kit 96T.

REF 5450601

LOT RR92B00.01

⌚ 2021-01-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 12	Adamts13 Ag Plate	RRS91B0	2021-04-30
1 x 0.3mL	Adamts13 Ag Conjugate	RQV91B0	2021-05-31
1x 80mL	Wash buffer Concentrate	RNA91B0	2021-03-31
1 x 90mL	Incubation buffer	RNB91B0	2021-03-31
1 x 12mL	TMB Substrate	RKN91B0	2021-01-31
1 x 12mL	Stop solution	RKK91B0	2021-02-28
1 x 0.5mL	Adamts13 Ag Calibrator 1	1RT81B1	2021-01-31
1 x 0.5mL	Adamts13 Ag Calibrator 2	2RT81B1	2021-01-31
1 x 0.5mL	Adamts13 Ag Calibrator 3	3RT81B1	2021-01-31
1 x 0.5mL	Adamts13 Ag Calibrator 4	4RT81B1	2021-01-31
1 x 0.5mL	Adamts13 Ag Calibrator 5	5RT81B1	2021-01-31
1 x 0.5mL	Adamts13 Ag High Control	RRY81B1	2021-01-31
1 x 0.5mL	Adamts13 Ag Low Control	RRZ81B1	2021-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Δ Cal 1 – Cal 5	≥ 0.7	1.35	OK
Verwendbarkeit Usability	High Control (RRY81B1)	0.53 – 0.89 IU/mL	0.76	OK
	Low Control (RRZ81B1)	0.12 – 0.28 IU/mL	0.19	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

26. JUNI 2019

Datum / Date:

Freigabe von
Approved by

Michael KAFKA *AK*

26. JUNI 2019

Datum / Date: