CERTIFICATE OF ANALYSIS

CHROMOGENIX Instrumentation Laboratory
A Werfen Company

PRODUCT NAME:	Antithrombin 1 x 25 IU	EXPIRATION DATE:	2021-07
PRODUCT NUMBER:	81079639		
LOT NUMBER:	N0797339		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION
	MIN	MAX	UNITS		PASS/FAIL
Antithrombin Activity	25	30	IU/vial	29	PASS
Assay Slope @ 37°C	-0.0024	-0.0034		-0.0029	PASS
Intercept @ 37°C	0.40	0.60		0.49	PASS
r^2	0.990	--		0.999	PASS

PREPARED BY: Maria Oettinger (PRINT) Maria Oettinger (SIGNATURE) DATE: 8-7-19

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz (PRINT) Debra Lee Harz (SIGNATURE) DATE: 8-12-19

TITLE: QC Technical Supervisor

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000