

CERTIFICATE OF ANALYSIS

CHROMOGENIX

Instrumentation
Laboratory
A Werfen Company

PRODUCT NAME:	<u>FXa, 10 x 71 nkat</u>	EXPIRATION DATE:	<u>2021-03</u>
PRODUCT NUMBER:	<u>82098539</u>		
LOT NUMBER:	<u>N1182397</u>		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION PASS/FAIL
	MIN	MAX			
FACTOR Xa, 71 nkat (650515)					
ACTIVITY	61	81	nkat/vial	71	PASS
STANDARD CURVE (r ²)	0.990	1.000	--	0.999	PASS
CONTROL RECOVERY	66.0	73.0	nkat/vial	67.9	PASS

PREPARED BY: Nehal Patel *N. M. Patel* DATE: 12/5/2018 12-5-18
 PRINT SIGNATURE

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Donna Tencza *Donna Tencza* DATE: 12/5/2018 12/5/18
 PRINT SIGNATURE

TITLE: Senior QC Analyst