

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 1	EXPIRATION DATE:	2021-07
PRODUCT NUMBER:	82265063		
LOT	N0797154		

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	DISPOSITION	
	MIN	MAX	UNITS			PASS/FAIL
Analysis with COATEST APC R:						
APT Time	CaCl ₂	28	34	seconds	32	PASS
	APC/CaCl ₂	90	--	seconds	121	PASS
APC Ratio		3.5	4.4	--	3.8	PASS

Analysis with COATEST APC R V:						
APT Time	CaCl ₂	--	38	seconds	35	PASS
	APC/CaCl ₂	75	--	seconds	108	PASS
APC Ratio		2.7	3.2	--	3.1	PASS

We hereby certify that the above product(s):

- have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
- meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Nehal Patel
PRINT

N. M. Patel
SIGNATURE

DATE: 7/19/2019

7-19-19

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz
PRINT

Debra Lee Harz
SIGNATURE

DATE: 7/19/2019

7/19/19

TITLE: Technical Supervisor