



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ2600.01

TECHNOVIEW® Arixtra CAL 5x1 mL

REF 5090010

LOT 2691B00.02

⌚ 2020-12-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1 mL	TECHNOVIEW Arixtra Cal 1	2181B00	2020-12-31
1x1 mL	TECHNOVIEW Arixtra Cal 2	2281B00	2020-12-31
1x1 mL	TECHNOVIEW Arixtra Cal 3	2381B00	2020-12-31
1x 1mL	TECHNOVIEW Arixtra Cal 4	2481B00	2020-12-31
1x1 mL	TECHNOVIEW Arixtra Cal 5	2581B00	2020-12-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Arixtra Cont low LOT 3U71B00	0.32 – 0.59 µg/mL	0.46	OK
	TECHNOVIEW Arixtra Cont high LOT 3V71B00	0.59 – 0.98 µg/mL	0.78	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK		

Durchführung von  
Prepared by

Suzana GRESAK-RODAKAR  
*Suzana Rodakar*

26. JUNI 2019

Datum / Date:

Freigabe von  
Approved by

*Eva Maria Hager*  
Eva Maria HAGER

26. JUNI 2019

Datum / Date: