

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ6PT00AZ
---	---	-----------

**TECHNOZYM<sup>®</sup> ADAMTS-13 INH Calibrator Set 5x0.5mL**

**REF**      5450461  
**LOT**      6PT91B0.02  
      2021-01-31

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT91B0	2021-01-31
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT91B0	2021-02-28
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT91B0	2021-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.54	OK
	Calibrator 5	< 0.1	0.02	OK
Verwendbarkeit Usability	High Control (RPY91B0)	56.55 – 94.25 U/mL	80.96	OK
	Low Control (RPZ91B0)	15.64 – 36.48 U/mL	29.57	OK

**Durchführung von  
Prepared by**

Fikret NASUFI *Fikret Nasufi*

11. SEP. 2019

Datum / Date:

**Freigabe von  
Approved by**

*Eva Maria Hager*  
Eva Maria HAGER

11. SEP. 2019

Datum / Date: