
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRQ00AZ
---	---	----------

TECHNOZYM[®] ADAMTS-13 ELISA Kit 2x48T.

REF 5450501

LOT RQ93B00.01

 2020-09-30

Inhalt Content	Reagenz Reagent	LOT	
2x48T	Adamts13 Plate	RQS91B0	2021-06-30
1x0.3mL	Adamts13 Conjugate	RQV71B1	2021-01-31
2x3mL	Adamts13 Activity Substrate	RQW81B0	2020-09-30
1x6mL	Adamts13 Antigen Substrate	RQX71B1	2021-01-31
1x0.7mL	Adamts13 Antigen Stable Peroxide	RQL71B1	2021-01-31
1x6mL	Adamts13 Antigen Stop Solution	RQK71B1	2021-01-31
1x80mL	Wash buffer Concentrate	RNA92B0	2021-06-30
1x90mL	Incubations buffer	RNB91B0	2021-03-31
1x0.5mL	Adamts 13 Calibrator 1	1RT81B1	2021-01-31
1x0.5mL	Adamts 13 Calibrator 2	2RT81B1	2021-01-31
1x0.5mL	Adamts 13 Calibrator 3	3RT81B1	2021-01-31
1x0.5mL	Adamts 13 Calibrator 4	4RT81B1	2021-01-31
1x0.5mL	Adamts 13 Calibrator 5	5RT81B1	2021-01-31
1x0.5mL	Adamts 13 High Control	RRY81B1	2021-01-31
1x0.5mL	Adamts 13 Low Control	RRZ81B1	2021-01-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High control RRY81B1 activity	0.44 – 0.74 IU/mL	0.52	OK
	High control RRY81B1 antigen	0.53 – 0.89 IU/mL	0.63	OK
	Low control RRZ81B1 activity	0.10 – 0.22 IU/mL	0.15	OK
	Low control RRZ81B1 antigen	0.11 – 0.25 IU/mL	0.20	OK
	Gain: 130 Activity ΔRFU/min Cal1-Cal5	≥200	339.8	OK
	Gain: 59 Antigen ΔRFU Cal1-Cal5	≥200	384	OK
Chargenspezifische Wertetabelle ist beigelegt / Batch specific table insert included		Ja / Yes	Ja / Yes	OK

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRQ00AZ
---	---	----------

**Durchführung von
Prepared by**

Fikret NASUFI *Fikret Nasufi*

13. AUG. 2019

Datum / Date:

**Freigabe von
Approved by**

Michael KAFKA *[Signature]*

13. AUG. 2019

Datum / Date: